

PRESSALIT.

Product	PLUS shower seat 450, manually height and manually sideways adjustable
Item number	R7454112
	VVS:777516212
	NRF:6317133
	EAN:5708590359922
	HMI: 124239
	FDA: INN/D135865
Colour	Anthracite grey
Product description	PLUS shower seats are offered in a number of colours. They are available as both fixed height, electrically adjustable and manually adjustable, with and without aperture, and with an option of back and armrests. They are suitable for wall-mounting or for fitting to Pressalit PLUS wall tracks.
	PLUS shower seat 450, with backrest. Manually height adjustable by 380 mm and sideways adjustable, without the use of tools. The seat is foldable. Materials: Comfortable polyurethane foam over a plywood core, fitted to an aluminiu profile. For fitting to horizontal PLUS wall track (not included). Replaces R7440
Adjustability	Height and sideways adjustable
Adjustment height	380 mm
Product capacity	Maximum user mass: 200 kg. When the seat is loaded with 200 kg, each wall fastener is loaded with 23 kg.
	The seat has been tested statically with safety factor 1.5, which is a requirement according to the applied assistive products standards.
Materials	The shower seat is made of polyurethane (PU) foam with a plywood core and this is covered with an in-mould coated surface that is protected using a clear PU lacquer. The backrest is made of polyurethane (PU) foam protected using a clear PU lacquer Internal components and fasteners are primarily made of stainless steel. The hinge barrel is capped with ABS.
Ergonomy	Ergonomically designed seat - contact surface are 450 x 437 mm. Soft and slip-resistant material that has a warm feel.



PRESSALIT®

Product Item number R7454112 The seat can fold up when not in use or to maximise floor space. The seat backrest can be adjusted simultaneously thanks to a safe and easy click sy Warranty Pressalit makes a commitment to remedy material and manufacturing defects period of five years from the invoicing date. Cleaning PLUS shower seat. The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use chlorine based, corrosive or abrasive cleaners as it may result is surface being damaged. The following cleaning agents have been tested by Pressalit and may be us long as they are diluted according to the instructions on the container. DISINFECTANTS: Erisan OXY+ TVättsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Bufa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven parts water. When the aluminum parts have been washed down with this solution must be wiped over with a damp cloth wrung out in clear water.	
The seat can fold up when not in use or to maximise floor space. The seat backrest can be adjusted simultaneously thanks to a safe and easy click sy Pressalit makes a commitment to remedy material and manufacturing defects period of five years from the invoicing date. PLUS shower seat. The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use chlorine based, corrosive or abrasive cleaners as it may result in surface being damaged. The following cleaning agents have been tested by Pressalit and may be use long as they are diluted according to the instructions on the container. DISINFECTANTS: Erisan OXY+ Tvätsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Bufa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven pare water. When the aluminum parts have been washed down with this solution	djustabl
Warranty Pressalit makes a commitment to remedy material and manufacturing defects period of five years from the invoicing date. PLUS shower seat. The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use chlorine based, corrosive or abrasive cleaners as it may result in surface being damaged. The following cleaning agents have been tested by Pressalit and may be us long as they are diluted according to the instructions on the container. DISINFECTANTS: Erisan OXY+ Tvättsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Büfa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven par water. When the aluminum parts have been washed down with this solution	
PLUS shower seat. The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use chlorine based, corrosive or abrasive cleaners as it may result in surface being damaged. The following cleaning agents have been tested by Pressalit and may be us long as they are diluted according to the instructions on the container. DISINFECTANTS: Erisan OXY+ Tvättsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Büfa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven par water. When the aluminum parts have been washed down with this solution	
PLUS shower seat. The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use chlorine based, corrosive or abrasive cleaners as it may result is surface being damaged. The following cleaning agents have been tested by Pressalit and may be us long as they are diluted according to the instructions on the container. DISINFECTANTS: Erisan OXY+ Tvättsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Büfa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained, surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven pai water. When the aluminum parts have been washed down with this solution	for a
long as they are diluted according to the instructions on the container. DISINFECTANTS: Erisan OXY+ Tvättsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Büfa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven par water. When the aluminum parts have been washed down with this solution	
Erisan OXY+ Tvättsprit 70% CaviCide surface disinfectant decontaminant cleaner Rodalon Desinfektionsspray Terralin protect (SCHÜLKE+) Dismonzon Plus (BODE Chemie GmbH) NON-DISINFECTANTS: Wetrok Caledor (Wetrok) Petrosol Citro (Büfa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven par water. When the aluminum parts have been washed down with this solution	sed, as
Wetrok Caledor (Wetrok) Petrosol Citro (Büfa) Important: Polyurethane surfaces of PLUS shower seats and PLUS backres contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven part water. When the aluminum parts have been washed down with this solution	
contains an antibacterial surface, which will be compromised if cleaned with other than soap and water. The product must be wiped down with a soft rag wrung out in clear water in after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven par water. When the aluminum parts have been washed down with this solution	
after cleaning, as some materials may otherwise be discolored and stained surface coating may deteriorate. DECALCIFICATION OF ALUMINUM PARTS: Pressalit recommends that chalky deposits be removed from wall tracks an aluminum parts using a solution of three parts 30% acetic acid to seven parts. When the aluminum parts have been washed down with this solution.	
Pressalit recommends that chalky deposits be removed from wall tracks and aluminum parts using a solution of three parts 30% acetic acid to seven parts. When the aluminum parts have been washed down with this solution	
	rts tap
Declaration of conformity CE MDR This product is in compliance with all relevant requirements of: - Fig. (EU) 2017/745 on medical devices.	≀egulatio



PRESSALIT.

Product	PLUS shower seat 450, manually height and manually sideways adjustable
Item number	R7454112

pressalit.com

Pressalit A/S
Pressalitvej 1
DK-8680 Ry
Tel.: +45 8788 8788
Fax: +45 8788 8789
E-mail: pressalit@pressalit.com