

## PRESSALIT.

	Shower mattress for MSCT 1
Product	
Item number	R844721304
	VVS:777542821
	EAN:5708590373935
	HMI: 128434
Colour	Pearl grey
Product description	Shower mattress for MSCT 1 shower change trolley, 660 x 2060 mm. Incl. 1.6 m hose.
	Shower mattress for MSCT 1 shower change trolley, 660 x 2060 mm. Incl. 1.6 m hose.
Materials	The mattress cover is welded polyester fabric with PVC-coating. It is UV-resistant and
Materiais	100% impermeable to water. It contains neither CFC nor latex. The foam filling is flexible polyether-based polyurethane foam, approved in accordance with Oeko-tex 100, Class I.
Warranty	Pressalit makes a commitment to remedy material and manufacturing defects for a period of three years from the invoicing date.



\_

## PRESSALIT.

Product	Shower mattress for MSCT 1
Item number	R844721304
Cleaning	<ul> <li>Mattress for MSCT.</li> <li>The product is to be cleaned with hot water or a mild soapy cleaning agent. Do not use chlorine based, corrosive or abrasive cleaners as it may result in the surface being damaged.</li> <li>The following cleaning agents have been tested by Pressalit and may be used, as long as they are diluted according to the instructions on the container.</li> <li>DISINFECTANTS: Sprit (husholdningssprit) Tvättsprit 70% Safe Zone Plus Selgiene Extreme Microsan Accel PREVention Wipes Lingettes</li> <li>The product must be wiped down with a soft rag wrung out in clear water immediately after cleaning, as some materials may otherwise be discolored and stained, and the</li> </ul>
Declaration of conformity	surface coating may deteriorate. MDR-ACC This product fulfils all relevant requirements in Regulation (EU) 2017/745 on medical devices in risk class I. This product is considered to be an <b>accessory</b> for a medical device, as it is not itself a medical device, but is intended to be used together with several particular medical devices to specifically enable the medical devices to be used in accordance with their intended purposes.

## pressalit.com

Pressalit A/S Pressalitvej 1 DK-8680 Ry Tel.: +45 8788 8788 Fax: +45 8788 8789 E-mail: pressalit@pressalit.com

